Corning Incorporated Life Sciences

Registered ISO 9001:2008

Product Description

Catalog Number: 4863

Product Description: Corning ® Universal Fit Pipette Tip, 1- 200 μL, 96 tip rack

Component Materials:

Tip - Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and

closures.

Rack - Polypropylene with orange colorant, meets *USP*, *Class VI* requirements for plastic

containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.

Lid - Polypropylene, meets *USP*. *Class VI* requirements for plastic containers and closures.

Product Dimensions:

Overall tip length - 1.995" I.D. small orifice - .028" Tolerences - +/-.050" I.D. large orifice - .215"

Special Feature:

The entire unit is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNse/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in ANSI/AAIM TS 72:2002/(R) 2010: Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing. Results: less than 0.1 EU/mL

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 3