

## Armenian Hamster IgG Isotype Control (eBio299Arm), eFluor 506, eBioscience™

| <b>Product Details</b> |                                     |
|------------------------|-------------------------------------|
| Size                   | 100 μg                              |
| Host/Isotype           | Armenian hamster / IgG              |
| Class                  | Control                             |
| Туре                   | Isotype Control                     |
| Clone                  | eBio299Arm                          |
| Conjugate              | eFluor® 506                         |
| Form                   | Liquid                              |
| Concentration          | 0.2 mg/mL                           |
| Purification           | Affinity chromatography             |
| Storage buffer         | PBS, pH 7.2, with 0.1% gelatin      |
| Contains               | 0.09% sodium azide                  |
| Storage conditions     | 4° C, store in dark, DO NOT FREEZE! |
| RRID                   | AB_2637310                          |

| Applications          | Tested Dilution | Publications |
|-----------------------|-----------------|--------------|
| Flow Cytometry (Flow) | Assay-Dependent | -            |
| Control (Ctrl)        | Assay-Dependent | -            |

## **Product Specific Information**

Description: The eBio299Arm monoclonal antibody is useful as an isotype control immunoglobulin.

Applications Reported: This eBio299Arm antibody has been reported for use in flow cytometric analysis.

Applications Tested: This eBio299Arm antibody has been tested by flow cytometric analysis of normal human peripheral blood cells and mouse splenocytes. Use the isotype control at the same concentration as the experimental antibody.

eFluor® 506 can be excited with the violet laser line (405 nm) and emits at 506 nm. We recommend using a 510/20 band pass filter, or equivalent. Please make sure that your instrument is capable of detecting this fluorochrome.

Excitation: 405 nm; Emission: 506 nm; Laser: Violet Laser.

Filtration: 0.2 µm post-manufacturing filtered.

For research use only, not for use in diagnostic procedures, not for resaile without express authorization. Products are warrantee to operate or perform susterstanding in comormance with published Product specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than the Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLED, ARE GRANTED INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to human or animals.