

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3495

Product Description: Costar ® Transwell-COL ®, 6.5mm, 0.4 µm pore size

Component Materials:

- | | | |
|----------------|---|--|
| Plate/Lid | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Transwell body | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Filter | - | Transparent collagen-treated PTFE membrane. |
| Collagen | - | Types I and III bovine placenta. |

Product Dimensions:

Length of Plate	-	5.030 in.	Bottom of Transwell to plate	-	.040 in.
Width of Plate	-	3.365 in.	Volume added / plate	-	0.6 mL
Height with Lid	-	.891 in.	Volume added / Transwell	-	0.1 mL
Tolerances of Dimensions	-	+/- .050 in.			

Sterilization:

The Transwells are sterilized according to ANSI/AAMI/ISO 11135 *Medical Devices - Validation and routine control of ethylene oxide sterilization* and released after successful sterility testing of Biological Indicators.

Sterility Assurance Level: SAL 10⁻⁶ The plates are irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Cell Attachment and Growth Characteristics:

The lot has been tested for the attribute of cell attachment and growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual and microscopic examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Collagen Assay:

Colorimetric determination of collagen attachment to the membrane.

Cell Attachment/Growth:

Four to five day cell attachment and growth to ≥ 95% confluency.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 6