

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 4808

**Product Description:** Costar® IsoTip Filter Tip, .5 - 10 µL 96 tip rack

**Component Materials:**

- Tip - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Rack - Polypropylene with white colorant, meets *USP, Class VI* requirements for plastic containers and closures.
- Lid - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

Overall tip length	-	2.085"	I.D. small orifice	-	.019"
Tolerances	-	+/- .050"	I.D. large orifice @ angle	-	.215"

**Sterilization:**

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10<sup>-3</sup>

**Special Feature:**

The entire unit is autoclavable.

**RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

**Pyrogens:**

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987)*.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 7