Performance Testing for Axygen[®] Automation Tip (PK-20-R)

Application Note



Method

The PerkinElmer Janus[®] liquid handling workstation was used to assess precision as coefficient of variation (% CV), and accuracy as percent deviation (% D) for Axygen 20 μ L tips.

To test the ability of the tip to dispense accurately and precisely at two dispense volumes, 2 μ L and 20 μ L, a rack of 96 tips aspirated from an Axygen low profile reservoir (Corning Cat. No. RES-SW96-LP) and dispensed into a Corning[®] 96-well, black, clear bottom microplate (Corning Cat. No. 3631).

For the 2 μL test volume, each tip aspirated 2 μL of Range C solution (Artel Cat. No. MVS-205) or DMSO Range C solution

(Artel Cat. No. MVS-217 solution) and dispensed 2 μ L into 198 μ L of diluent solution (Artel Cat. No. MVS-202) in each well. For the 20 μ L test volume, each tip aspirated 20 μ L of Range B solution (Artel Cat. No. MVS-204) and dispensed 20 μ L into 180 μ L of diluent solution in each well. To determine the volume of liquid dispensed in each well, absorbance readings for the solutions (diluted Range C solution for 2 μ L dispense and Range B solution for 20 μ L dispense) were measured using an Artel ELx800NB[®] plate reader (Artel Cat. No. 1311197). Each study was performed 3 independent times for a total of 288 tip dispenses. Evaluation criteria include % D from the set dispense volume and % CV of the measured dispenses.

Results




Figure 1. Analysis of PK-20-R tip with aqueous dispense. The precision (assessed by % CV) and accuracy (assessed by % D) of Axygen PK-20-R tips dispensing (A) 2 μ L and (B) 20 μ L volumes using the PerkinElmer Janus liquid handling workstation were determined using the Artel MVS[®] system. The % CV and % D were below 1.5 % for both 2 μ L and 20 μ L dispenses, n = 288.

Table 1. Aqueous Dispense Results

Target Volume (μL)	2	20
n	288	288
% CV	1.32± 0.19	1.04 ± 0.21
% D	0.53 ± 0.30	0.42 ± 0.08
Outliers	0	0

Conclusion

The % CV and % D for the Axygen automation PK-20-R tips dispensing 2 μ L and 20 μ L were 5% or below. Therefore, Axygen automation PK-20-R tips can precisely and accurately dispense volumes as low as 2 μ L and as high as 20 μ L for aqueous and DMSO solutions using the PerkinElmer Janus liquid handling workstation.




Figure 2. Analysis of PK-20-R tip with DMSO dispense. The precision (assessed by % CV) and accuracy (assessed by % D) of Axygen PK-20-R dispensing 2 μ L volumes using the PerkinElmer Janus[®] liquid handling workstation were determined using the Artel MVS[®] system. The % CV and % D were below 1% for the 2 μ L dispense, n = 288.

Table 2. DMSO Dispense Results

Target Volume (μL)	2
n	288
% CV	0.99 ± 0.15
% D	0.26 ± 0.21
Outliers	0



www.corning.com/lifesciences/solutions

In our continuous efforts to improve efficiencies and develop new tools and technologies for life science researchers, we have scientists working in Corning R&D labs doing what you do every day, across the globe. From collection to analysis, our technical experts understand your challenges and your need for simplified efficient, low- to high-throughput genomics processes.

A combination of global manufacturing expertise, extensive use of in-house automation, an unsurpassed commitment to product innovation and a thorough understanding of your processes enables Corning to offer a beginning-to-end portfolio of high-quality, reliable consumables and reagents for genomics applications.

For more specific information on claims, visit the Certificates page at www.corning.com/lifesciences.

Warranty/Disclaimer: Unless otherwise specified, all products are for research use only. Not intended for use in diagnostic or therapeutic procedures. Not for use in humans. Corning Life Sciences makes no claims regarding the performance of these products for clinical or diagnostic applications.

GOSSELIN

PYREX

Life Sciences 836 North St. Building 300, Suite 3401 Tewksbury, MA 01876 t 800.492.1110 t 978.442.2200 f 978.442.2476 www.corning.com/lifesciences

Corning Incorporated

CORNING | FALCON AXYGEN

For a listing of trademarks, visit www.corning.com/clstrademarks. All other trademarks are the property of their respective owners.