

Performance Testing for Axygen® Automation Tip (RRP-50-CBK-R)

Application Note



Method

The Varispan™ arm on the Perkin Elmer Janus® liquid handling workstation was used to assess precision as coefficient of variation (% CV), and accuracy as percent deviation (% D), for Axygen 50 µL tips.

To test the ability of the tip to dispense accurately and precisely at two dispense volumes, 5 µL and 50 µL, a column of 8 tips aspirated from an Axygen low profile reservoir (Corning Cat. No. RES-SW96-LP) and dispensed into a Corning® 96-well black/clear-bottom microplate (Corning Cat. No. 3631).

For the 5 µL test volume, each tip aspirated 5 µL of Range C solution (Artel Cat. No. MVS-205) or DMSO Range C solution

(Artel Cat. No. MVS-217 solution) and dispensed 5 µL into 195 µL of diluent solution (Artel Cat. No. MVS-202) in each well. For the 50 µL test volume, each tip aspirated 50 µL of Range B solution (Artel Cat. No. MVS-204) and dispensed 50 µL into 150 µL of diluent solution in each well. To determine the volume of liquid dispensed in each well, absorbance readings for the solutions (diluted Range C solution for 5 µL dispense and Range B solution for 50 µL dispense) were measured using an Artel ELx800NB® plate reader (Artel Cat. No. 1311197). Each study was performed 6 independent times for a total of 48 tip dispenses. Evaluation criteria include % D from the set dispense volume and % CV of the measured dispense volume for the 48 tip dispenses.

Results




Figure 1. Analysis of RRP-50-CBK-R tip with aqueous dispense. The precision (assessed by % CV) and accuracy (assessed by % D) of Axygen RRP-50-CBK-R tips dispensing (A) 5 µL and (B) 50 µL volumes using the Varispan arm on the Perkin Elmer Janus liquid handling workstation were determined using the Artel MVS® system. The % CV and % D were below 3% for both 5 µL and 50 µL dispenses, n = 48.

Table 1. Aqueous Dispense Results

Target Volume (µL)	5	50
n	48	48
% CV	1.35 ± 0.21	0.28 ± 0.13
% D	2.0 ± 0.66	1.19 ± 0.07
Outliers	0	0

Conclusion

The % CV and % D for the Axygen automation RRP-50-CBK-R tips dispensing 5 µL and 50 µL were 5% or below. Therefore, Axygen automation RRP-50-CBK-R tips can precisely and accurately dispense volumes as low as 5 µL and as high as 50 µL for aqueous and DMSO solutions using the Varispan arm on the Perkin Elmer Janus liquid handling workstation.





Figure 2. Analysis of RRP-50-CBK-R tip with DMSO dispense. The precision (assessed by % CV) and accuracy (assessed by % D) of Axygen RRP-50-CBK-R tips dispensing 5 µL volumes using the Varispan arm on the Perkin Elmer Janus liquid handling workstation were determined using the Artel MVS system. The % CV and % D were below 1% for the 5 µL dispense, n = 48.

Table 2. DMSO Dispense Results

Target Volume (µL)	5
n	48
% CV	0.50 ± 0.16
% D	0.48 ± 0.18
Outliers	0



www.corning.com/lifesciences/solutions

In our continuous efforts to improve efficiencies and develop new tools and technologies for life science researchers, we have scientists working in Corning R&D labs doing what you do every day, across the globe. From collection to analysis, our technical experts understand your challenges and your need for simplified efficient, low- to high-throughput genomics processes.

A combination of global manufacturing expertise, extensive use of in-house automation, an unsurpassed commitment to product innovation and a thorough understanding of your processes enables Corning to offer a beginning-to-end portfolio of high-quality, reliable consumables and reagents for genomics applications.

For more specific information on claims, visit the Certificates page at www.corning.com/lifesciences.

Warranty/Disclaimer: Unless otherwise specified, all products are for research use only. Not intended for use in diagnostic or therapeutic procedures. Not for use in humans. Corning Life Sciences makes no claims regarding the performance of these products for clinical or diagnostic applications.

For additional product or technical information, visit www.corning.com/lifesciences or call 800.492.1110. Outside the United States, call +1.978.442.2200 or contact your local Corning sales office.

Corning Incorporated
Life Sciences

836 North St.
Building 300, Suite 3401
Tewksbury, MA 01876
t 800.492.1110
t 978.442.2200
f 978.442.2476

www.corning.com/lifesciences

CORNING | **FALCON** | **AXYGEN** | **GOSSSELIN** | **PYREX**

For a listing of trademarks, visit www.corning.com/clstrademarks.
All other trademarks are the property of their respective owners.