Corning Inc Life Sc <u>Regist</u> ISO 900	ences ered Product Description
Catalog Number:	3407
Product Description:	Costar ® Snapwell TM, 12mm, TC treated, 0.4 $\mu m$ pore size, ina 6-well cluster plate with lid
<b>Component Materials:</b> Plate/Lid Body & Ring Filter	<ul> <li>Virgin Polystyrene, meets USP, Class IV requirements for plastic containers and closures.</li> <li>Virgin Butadiene/Styrene Block Co-polymer</li> <li>Translucent polycarbonate membrane, meets USP, Class VI requirements for plastic containers and closures.</li> </ul>
<b>Product Dimensions:</b> Length of Plate Width of Plate Tolerances of Dimensions	<ul> <li>5.030 in. Height with Lid891 in.</li> <li>3.365 in. Bottom of Snapwell to060 in plate</li> <li>+/050 in.</li> </ul>

# Special Feature:

The bottom section of the Snapwell is easily detachable from its upper assembly for placement in the Diffusion Chamber.

### Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10<sup>-3</sup>

## Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

## Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Cell Culture Treatment: Wettability test using water to insure the presence of a hydrophilic

	surface.
Visual Attributes:	Visual and microscopic examination of the product.
Packaging:	Inspection for seal and barrier integrity, accurate labeling, and correct
	product configuration.

#### Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5