

Quick and easy dispenser function check!



Testing Flasks BLAUBRAND® for Dispensers

F I R S T C L A S S · B R A N D

Using the **BRAND testing flask with 3 marks**, you can immediately see if your bottle-top dispenser is working correctly.

- Class A, DAkkS-calibrated, with certificate
- Quick and easy usage
- For use with common bottle-top dispensers
- Available for dispensers from 10 ml to 100 ml




Dispense with confidence!

Malfunction of bottle-top dispensers, such as a case in which residues cause valves to stick, can cause significant volume deviations and consequently wrong results. BRAND testing flasks with 3 volume marks let you make a quick function check of your bottle-top dispenser.

The mark in the middle shows the nominal volume; the upper and lower marks show the specified error limits ($EL = A + 2 CV$ for $A \leq \pm 0.5 \%$ and $CV < 0.1 \%$). **If the meniscus of the dispensed nominal volume falls between the upper and lower error limits, the dispenser is functioning correctly.**

If the dispensed volume exceeds the error limits, we recommend careful cleaning of the instrument, or, if necessary, following of the troubleshooting procedure described in the operating manual. Then, perform a gravimetric volume test as per the manufacturer's instructions. If this gravimetric volume test also returns results outside of the error limits, please contact the manufacturer. The dispenser may need to be sent in for repair.

Volumetric Flask, DAkkS calibrated



Note:

This quick function test is not intended to replace the gravimetric test of the monitoring of measuring instruments according to DIN EN ISO 8655.



Technical Data

Volumetric flasks with 3 marks, DAkkS calibrated

BLAUBRAND®, class A, DE-M marking

Boro 3.3, DIN EN ISO 1042. Calibrated to contain (TC, In). Incl. DAkkS certificate. Pack of 1. **Delivery date to be agreed upon.**

| Capacity ml | Upper/lower mark ± ml | Neck inner Ø mm | Neck size | With PP stopper Cat. No. |
|-------------|-----------------------|-----------------|-----------|--------------------------|
| 10 | 0.070 | 7 ± 1 | 10/19* | 382 04 |
| 25 | 0.175 | 9 ± 1 | 10/19 | 382 06 |
| 50 | 0.35 | 11 ± 1 | 12/21 | 382 08 |
| 100 | 0.70 | 13 ± 1 | 14/23 | 382 10 |

* enlarged ground joint

Further sizes on request.

For more information on the BLAUBRAND® volumetric instruments, please visit www.brand.de!

BLAUBRAND®, BRAND®, BRAND. For lab. For life.® and the BRAND word and figurative mark are registered trademarks of BRAND GMBH + CO KG, Germany.

Other reproduced brands are the property of the respective owner.

Our technical literature is intended to inform and advise our customers. However, the validity of general empirical values, and of results obtained under test conditions, for specific applications depends on many factors beyond our control. Please appreciate, therefore, that no claims can be derived from our advice. The user is responsible for checking the appropriateness of the product for any particular application.

California Residents: For more information concerning California Proposition 65, please refer to www.brand.de/calprop65

Subject to technical modification without notice. Errors excepted.

BRAND GMBH + CO KG · P.O. Box 11 55 · 97861 Wertheim · Germany
Tel.: +49 9342 808-0 · Fax: +49 9342 808-98000 · E-Mail: info@brand.de · Internet: www.brand.de

