

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 4133

**Product Description:** DeckWorks™ Pipet Tips, 1-300 µL, Graduated, Polypropylene, Natural, Not Sterile, Reload Starter Kit, 96 Tips/Deck, 10 Decks/Reload Pack (960 Tips), 1 Reload Pack plus 10 Empty Hinged Racks/Case (960 Tips)

**Component Materials:**

- Tip - Polypropylene, virgin, color natural.
- Rack - Polypropylene, lid color natural, base color orange. Heavy metal free color concentrate (meets CONEG requirements).
- Deck - Polypropylene, color gray. Heavy metal free color concentrate (meets CONEG requirements).
- Base - PET, color natural.
- Lid - PET, color natural.

**Product Dimensions:**

- Tip: Overall Length 58.6 mm, I.D. (small orifice) 0.51 mm.
- Deck: 96 holes in 8 x 12 pattern on 9 mm centers.
- Rack: Length 142 mm x Width 110 mm x Height 69 mm.
- Reload Pack: Length 231 mm x Width 165 mm x Height 169 mm.

**Sterilization:**

Supplied Not Sterile.

**RNase/DNase Testing:**

This product is tested to confirm that no detectable RNase/DNase contamination is present.

**Pyrogens:**

This product is tested per the criteria established in *Guideline on Validation of the Limulus Amebocyte Lystate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices* (FDA, December, 1987). Results less than 0.1 EU/mL.

**PCR Certification:**

This product is tested to confirm that no detectable PCR inhibitors are present.

**Autoclaving:**

DeckWorks™ tips, decks, hinged racks, station layer trays, station covers and station bases may be autoclaved using a standard cycle of 121°C and 15 psi for 15 minutes. Place tips into deck/rack or into station trays before autoclaving.

**Lot Number Designation:**

Each production lot of this product is uniquely identified with an 8 digit lot number: First 3 digits - Julian date; Next 2 digits - last 2 digits of year of manufacture; Last 3 digits - Batch identification.

Rev No: 1