Corning Incorporated Life Sciences

Registered

Product Description

Catalog Number: 3320

Product Description: Corning ® CellSTACK® - 10 Chamber with CellBIND® surface

Component Materials:

Top Plate - Virgin Polystyrene, meets USP Class VI requirements for plastic containers and

closures.

Middle Plate - Virgin Polystyrene, meets USP Class VI requirements for plastic containers and

Closures.

Bottom Plate - Virgin Polystyrene, meets USP Class VI requirements for plastic containers and

closures.

Adhesive - Proprietary Acrylate, meets ISO-10993, Biocompatibility requirements and does

not contain any animal products.

Vent Cap - Virgin High-Density Polyethylene (HDPE), meets USP, Class VI requirements

for plastic containers and closures, with heavy metal-free (meets CONEG requirements) color concentrate; 0.2 µm pore size, modified, non-woven polytetrafluoroethylene (PTFE) membrane meeting USP, Class VI requirements

for plastic containers and closures.

Product Dimensions:

 Overall Length
 13.2 in.(335mm)
 Overall Width
 8.1 in. (206mm)

 Overall Height with cap
 8.0 in. (203mm)
 Neck ID
 1.0 in. (26mm)

 Tolerances
 +/-0.1 in. (2.5mm)
 Neck O.D. incl. Threads
 1.3 in. (32mm)

Distance between plates - 0.67 in. (17mm)

Total Cell Growth Area - 6360 cm²

Recommended Working Volume - 1300 - 2000mL

Animal Content - Product does not contain materials of animal origin

Non-Pyrogenic – Vessel tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is ≤ 0.10 EU/ml or ≤ 4 EU/device.

Sterilization - Product has been sterilized and dosimetrically released per the requirements ANSI/AAMI/ISO 11137 "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10⁻⁶.

Sterility - Products labeled Sterile Fluid Path have been designed to ensure sterility of the portion of the product intended for contact with fluids.

Tissue Culture - Tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line. A minimum of 95% confluency is required for acceptance.

Surface Characterization – Surface is characterized to be hydrophilic, negatively charged and composed of ≥20% oxygen atoms. The surface composition has been developed for cell attachment and growth

Corning Incorporated Life Sciences

Registered ISO 9001

Product Description

Catalog Number: 3320

Product Description: Corning ® CellSTACK® - 10 Chamber with CellBIND® surface

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Packaging Inspection - Pass
Cell Attachment & Growth Treatment Verification – Pass
Leak Test – Pass

Lot Number Designation - 8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.

Serial Number Designation - 12 Digit Serial Number: First 8 digits – Lot number (see lot number designation above); Next 3 digits CellSTACK serial number; Last digit – Stack designation

Rev 16