

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 4807

Product Description: Corning® IsoTip Filter Tip, .2 - 10 µL, 96 tip rack

Component Materials:

- Tip - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Rack - Polypropylene with orange colorant, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.
- Lid - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- Overall tip length - 1.300" I.D. small orifice - .015"
- Tolerances - +/- .050"

Sterilization:

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Special Feature:

The entire unit is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987)*. Results: less than 0.1 EU/mL

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification. The entire lot is sterilized as one unit.

Rev No: 5