

Catalog Number: 11705

Product Description: Corning® 50mL centrifuge tube, self-standing, polypropylene, with accessories and dip tube, Non-pyrogenic

Component Materials:

- Centrifuge Tube - Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures.
- Cap - Virgin High Density Polyethylene meets USP Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.

Accessories:

- Tubing - Chemically resistant, heat sealable, Thermoplastic Elastomer, meets USP, Class VI requirements for plastic containers and closures.
- Connector - Polypropylene, meets USP, Class VI requirements for plastic containers and closures.
- Filter - 25mm/0.2µm Acrylic Proprietary formulation, meets USP, Class VI requirements for Biological Test for plastics.
- dip tube - Polypropylene, meets USP, Class VI requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|------------------------|---|----------|----------------|---|-------------|
| Length of tube w/cap | - | 5.07 in. | Length of Tube | - | 4.52 in. |
| Outer Diameter of tube | - | 1.24 in. | Tolerances | - | +/-0.05 in. |

BSE/TSE – Product complies with the latest revision of EMA/410/01 “Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products” by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Non-Pyrogenic – Tested and met the criteria established in the current version of ANSI/AAMI ST 72, “Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing” and USP <85>, “Bacterial Endotoxins Test”. The acceptance level for product is ≤ 0.5 EU/mL or ≤ 20 EU/device.

Sterility – Products labeled Sterile Fluid Path have been designed to ensure sterility of the portion of the product intended for contact with fluids.

Sterilization – Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, “Sterilization of health care products – Radiation”. Products meet a minimum Sterility Assurance Level (SAL) of 10^{-6} .

Quality Control Testing – Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection – Pass
- Packaging Inspection – Pass
- Integrity Test – Pass
- Centrifugation Test – Pass

Lot Number Designation -

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.