## Corning Incorporated Life Sciences

Registered ISO 9001:2008

## **Product Description**

Catalog Number: 3402

Product Description: Costar ® Transwell ®, 12mm, TC treated, 3.0 µm pore size, in a 12-well cluster plate

with lid

**Component Materials:** 

Plate/Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and

closures.

Transwell body - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Filter - Translucent polycarbonate membrane, meets *USP*, *Class VI* requirements for plastic

containers and closures.

**Product Dimensions:** 

Length of Plate - 5.035 in. Bottom of Transwell to - .040 in.

plate

Width of Plate - 3.370 in. Volume added / plate - 1.5 mL Height with Lid - .890 in. Volume added / - 0.5 mL

Transwell

Tolerances of - +/-.050 in.

**Dimensions** 

## Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10<sup>-3</sup>

## Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Cell Culture Treatment: Wettability test using water to insure the presence of a hydrophilic

surface.

Visual Attributes: Visual and microscopic examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct

product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5