

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 7079-2N

**Product Description:** PYREX® Pipet, Disposable Glass, Serological, 2 mL in 1/100 mL Bulk Pack, Non-Sterile, Not Plugged

**Component Materials:**

Type I Class B - Alumino-Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

**Attributes:**

ASTM Reference - E-714 (Specification for Disposable Glass Serological Pipets)  
Capacity - 2.0 mL ± 0.06 mL at any mark (3% of full capacity)  
Product Dimensions - 6 mm OD x 274 mm Long

**Autoclaving:**

This is a disposable product and is not intended for reesterilization. It will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

**Heating / Cooling:**

This product is primarily designed for a specific use that does not involve direct heating. In normal use, temperatures will be from -50° to +125°C. This is below the temperature at which thermal failure could be a problem.

**Vacuum / Pressure:**

We DO NOT test this product for vacuum and therefore cannot recommend vacuum greater than 50 mm or Mercury (1 psi negative pressure).

We DO NOT test this product for pressure and therefore cannot recommend pressure greater than 100 mm of Mercury (2 psi).

**Sterilization:**

Shipped Not Sterile

**TSE/BSE:**

Materials of animal origin are not used in the manufacture of this product.

**Lot Number Designation:**

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Revision Date:**

9 OCT 2013

Rev No: 2