# Corning Incorporated Life Sciences

Registered ISO 9001:2008

## **Product Description**

Catalog Number: 4872

Product Description: Costar ® Reagent reservoir, 100 mL

**Component Materials:** 

Reservoir - Virgin white polystyrene

**Product Dimensions:** 

**Dimensions** 

#### Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10<sup>-3</sup>

## Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct

product configuration.

### Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification. The entire lot is sterilized as one unit.

Rev No: 4