

Corning Incorporated  
Life Sciences

Registered ISO 9001

**Catalog Number** LC2000-26  
**PDS reference** PDS\_Cbor\_5079 Version \_ 1  
**Product description** Corning® Gosselin™ Square HDPE Bottle, 2 L, Graduated with 24H Urines Label, 58 mm Red Cap with Seal, Non-assembled, 50/C

**Sanitary Assurance Level**

Non sterile

**Expiration Statement**

N/A

**Performance testing (AQL)**

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

**Primary body composition**

HDPE

**Secondary body composition**

HDPE

**Closure**

58 mm Red Cap with Seal

**Special feature**

CE marked product, Non-assembled, w

**Packaging**

50/Case

## Regulation Statement

We hereby confirm that we and/or our raw materials suppliers do not intentionally use the following substances in the manufacturing of the product:

- LATEX
- BISPHENOL A (CAS No 80-05-7)
- PHTHALATES as per Regulation 1907/2006/EC (SVHC).

Raw material suppliers also guarantee the following requirements through certificates of compliance:

- EU Food Contact 1935/2004/EC and 10/2011/EU.  
However, our finished products have not been validated for food contact applications. They are designed for laboratory applications only.
- BSE/TSE compliant: the animal origin materials and/ or derived materials were not intentionally used during manufacturing process.  
As such, our finished products do not contain substances considered at risk for transmitting BSE/TSE.
- Complies to REACH SVHC candidate list (Regulation 1907/2006/EC). Do not contain more than 0.1% (w/w) of any substance of very high concern listed in the ECHA candidate list in accordance with Article 33 of the Reach Regulation.
- Heavy Metals as per directive 94/62/EC: meets the relevant requirements on packaging and packaging waste.  
The sum of the concentration Cd, Cr (VI), Hg and Pb does not exceed 100 ppm.



**Hugues Vanderschelden**  
Plant Quality Manager, Borre  
vandersch2@corning.com