# Corning Incorporated Life Sciences

Registered ISO 9001:2008

# **Product Description**

Catalog Number:

70800-15

Product Description:

PYREX® VISTA™ Tube, Test, Beaded Rim, 15 x 125 mm

# Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

# Attributes:

Capacity - 14 mL Product Dimensions - 15 mm OD x 125 mm High, 1.0-1.2 mm Wall

#### Autoclaving:

This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

#### Heating / Cooling:

This product is designed for heating by hot plate, flame, oven, microwave or heating mantles. In normal service it can be heated many times to 230°C. It can be heated one time to 490°C in extreme service.

This product can be cooled to -230°C. DO NOT plunge product directly into a liquid coolant. Place the product in the vapor phase of the liquified gasses for slow cooling prior to immersion in the coolant.

#### Vacuum / Pressure:

This product is not designed for and is not rated for vacuum or pressure applications.

#### Sterilization:

Shipped Not Sterile

### TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

# Lot Number Designation:

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

# **Revision Date:**

11 MAR 2013

Rev No: 1