

Pierce Agarose ChIP Kit

MAN0016149

Pub. Part No. 2162216.9

Rev A.0

26156

Number	Description
26156	<p>Pierce Agarose ChIP Kit, contains sufficient reagents to perform ChIP reactions for 30 target proteins from 10 sample conditions</p> <p>Kit Contents:</p> <p>ChIP Grade Protein A/G Plus Agarose, 0.65mL; store at 4°C</p> <p>IP Dilution/Wash Buffer (5X), 13mL, store at 4°C</p> <p>IP Wash Buffer 3 (5X), 4.5mL; store at 4°C</p> <p>IP Elution Buffer (2X), 4.5mL; store at 4°C</p> <p>Sodium Chloride (5M), 3mL; store at 4°C</p> <p>Column Accessory Pack, 3 packs, 10 spin columns, 20 collection tubes and 10 plugs; store at room temperature or 4°C</p> <p>Microcentrifuge Tubes, 1.5mL, 75 tubes; store at room temperature or 4°C</p> <p>DNA Clean-Up Column and Reagents, 40 purifications; store at room temperature or 4°C</p> <ul style="list-style-type: none">DNA Clean-Up Column, 40 columnsDNA Column Binding Solution, 30mLDNA Column Wash Solution, 6mLpH Indicator, 0.8mLDNA Column Elution Solution, 5mL <p>Anti-RNA Polymerase II Antibody, 25µL; store at -20°C</p> <p>Normal Rabbit IgG, (1mg/mL), 10µL; store at -20°C</p> <p>ChIP Positive Control Primers (GAPDH promoter; human specific), 100µL; store at -20°C</p> <p><u>Pierce Chromatin Prep Module (Product No. 26158)</u></p> <p>Membrane Extraction Buffer, 10mL; store at 4°C</p> <p>Nuclear Extraction Buffer, 10mL; store at 4°C</p> <p>MNase Digestion Buffer, 5mL; store at 4°C</p> <p>MNase Stop Solution, 0.5mL; store at 4°C</p> <p>Halt Protease and Phosphatase Inhibitor Cocktail EDTA-free (100X), 4 × 90µL microtubes; store at 4°C</p> <p>Glycine Solution (10X), 15mL; store at room temperature</p> <p>PBS (20X), 15mL; store at 4°C</p> <p>Micrococcal Nuclease (ChIP Grade) (10 U/µL), 25µL; store at -20°C</p> <p>Proteinase K (20mg/mL), 0.25mL; store at -20°C</p> <p>DTT (7.7mg), Lyophilized, 1 vial; store at room temperature or 4°C</p> <p>Storage: Upon receipt store contents of yellow pouch at -20°C. Store all other components as indicated. The kit is shipped at ambient temperature.</p>

Introduction

The Thermo Scientific™ Pierce™ Agarose ChIP Kit provides a simple, fast and reproducible method to perform chromatin immunoprecipitation (ChIP) assays. ChIP assays identify links between the genome and the proteome by monitoring transcription regulation through histone modification¹ (epigenetics) or transcription factor-DNA binding interactions. The strength of the ChIP assay technique is its ability to capture a snapshot of specific protein-DNA interactions as they occur in living cells and then quantitate the interactions using standard or quantitative PCR.^{2,3}

The kit contains all reagents to perform a successful ChIP assay. The specially titrated and tested micrococcal nuclease digests the DNA, eliminating variable results caused by the traditional method of sonication. The advantages of enzymatic digestion include reproducibility of digestion, control of the reaction, and easy titration of the enzyme for each specific cell type. The specially blocked ChIP Grade Pierce Protein A/G Plus Agarose Resin provides high-binding capacity and low background. Performing the ChIP, wash, and elution steps in the included spin columns is fast, convenient and minimizes sample loss. Using these columns and optimized reagents, the Pierce ChIP Kit protocol is streamlined to achieve complete crosslink reversal, protein digestion, and DNA purification with minimal time and sample handling.

Additional Materials Required

- Mammalian cell culture reagents and equipment
- Chemical fume hood
- ChIP-qualified primary antibody against the DNA binding protein of interest
- Cell scrapers
- Thermomixer or heat block
- 50 and 15mL conical tubes
- 16% Formaldehyde (Product No. 28906)
- DMSO (optional)
- Nuclease-free water
- Ethanol (95-100%)
- Standard or quantitative PCR reagents and equipment
- Oligonucleotide primers specific to the gene promoter of interest

Protocol for ChIP Assay

Note: Read the entire protocol before beginning the assay.

A. Experimental Design

- When designing a ChIP experiment, consider the number of immunoprecipitations desired from the chromatin sample and the number of cell culture conditions (e.g., drug-treated vs. non-treated). For accurate comparison, immunoprecipitate the same target proteins and controls from each cell culture condition.
- Performing a ChIP using the Normal Rabbit IgG is an effective negative control.
- The included anti-RNA polymerase II antibody and GAPDH primers provide a positive control for assay technique and reagent integrity. Perform a single RNA polymerase II immunoprecipitation on isolated chromatin from the control (non-treated) cell culture condition. Amplify the resulting enriched DNA with the provided GAPDH primer pair. The strength of the RNA Polymerase II IP signal should be > 50% relative to the total input control in a well-functioning assay (see Troubleshooting Section).
- If performing multiple ChIPs from a single cell culture condition, crosslinked chromatin may be prepared in bulk. Scale the reagent amounts according to the number of ChIPs desired (Table 1) and sub-divide the chromatin sample after diluting with the 1X IP Dilution Buffer. The total input control obtained from the cell culture condition is common to this set of ChIPs.

Table 1. Reagent volumes to use based on the number of ChIPs.

<u>Example number of HeLa cells</u>	<u>2 × 10⁶</u>	<u>6 × 10⁶</u>	<u>2 × 10⁷</u>	<u>6 × 10⁷</u>
Number of ChIPs	1	3	10	30
1X PBS	2.1mL	6.3mL	21mL	63mL
Lysis Buffer 1	0.1mL	0.3mL	1mL	3mL
Lysis Buffer 2	0.05mL	0.15mL	0.5mL	1.5mL
MNase Digestion Buffer Working Solution	0.1mL	0.3mL	1mL	3mL
1X IP Dilution Buffer	0.45mL	1.2mL	4.5mL	12mL
IP Wash Buffer 1	0.5mL	1.5mL	5mL	15mL
IP Wash Buffer 2	1mL	3mL	10mL	30mL
IP Wash Buffer 3	0.5mL	1.5mL	5mL	15mL
1X IP Elution Buffer	0.15mL	0.45mL	1.5mL	4.5mL

B. Material Preparation

- Prepare the reagents listed below once and use for all 30 ChIP reactions.

Glycine Solution (10X)	If precipitate formed, warm solution to 37°C in a water bath for 30 minutes and vortex before use.
PBS	If precipitate formed, warm solution to 37°C in a water bath for 30 minutes and vortex before use. Dilute PBS (20X) to 1X with nuclease-free water.
1M DTT	Add 50µL of ultrapure water to the DTT. Store this solution at -20°C for up to 6 months.
DNA Column Binding Buffer	Add 120µL of pH Indicator Solution to the vial of 30mL DNA Column Binding Solution and mix well. Store at room temperature.
DNA Column Wash Buffer	Add 24mL of ethanol (95-100%) to the vial of 6mL DNA Column Wash Solution provided and mix well. Store at room temperature.

- The amounts listed below are for preparing one ChIP. If you are preparing chromatin in bulk or performing multiple ChIPs, multiply the reagent amounts by the number of ChIPs being performed (see Table 1).

Lysis Buffer 1	Add 1µL of the Halt Cocktail to 100µL of Membrane Extraction Buffer in a microcentrifuge tube and place on ice.
Lysis Buffer 2	Add 0.5µL of the Halt Cocktail to 50µL of Nuclear Extraction Buffer in a microcentrifuge tube and place on ice.
MNase Digestion Buffer Working Solution	Add 0.1µL of 1M DTT to 100µL of MNase Digestion Buffer and place tube at room temperature.
1X IP Dilution Buffer	Add 100µL of IP Dilution/Wash Buffer (5X) and 5µL of Halt Cocktail to 395µL of nuclease-free water. Store at 4°C.
IP Wash Buffer 1	Prepare 0.5mL by combining 0.1mL IP Dilution/Wash Buffer (5X) with 0.4mL nuclease-free water. Store at 4°C.
IP Wash Buffer 2	Prepare 1mL by combining 0.2mL IP Dilution/Wash Buffer (5X) and 70µL Sodium Chloride (5M) with 0.73mL of nuclease-free water. Store at 4°C.
1X IP Wash Buffer 3	Prepare 0.5mL by combining 0.1mL of IP Wash Buffer 3 (5X) with 0.4mL of nuclease-free water. Store at 4°C.
1X IP Elution Buffer	Warm the IP Elution Buffer (2X) in a 37°C water bath until fully dissolved. For each IP and total input control, prepare 150µL of 1X IP Elution Buffer by combining 75µL of IP Elution Buffer (2X) with 75µL nuclease-free water. Store at room temperature.

C. Crosslinking and Cell Pellet Isolation

Note 1: Perform all centrifugations at room temperature.

Note 2: This procedure is for one CHIP. Multiply reagent amounts by the number of CHIPs being performed (see Table 1).

1. Culture adherent mammalian cells and treat as desired.

Optional: If you are unfamiliar with cell type being used, culture an extra dish of cells for determining cell number. Before crosslinking, trypsinize and determine the cell number from the extra dish of cells.

2. To each dish containing cell culture media, add sufficient quantity of 16% formaldehyde to obtain a final concentration of 1% formaldehyde.
Caution: Formaldehyde is a skin irritant and the fumes are toxic. Use proper personal protective, laboratory safety and disposal equipment.
3. Mix well by gently swirling the dish. Incubate at room temperature for 10 minutes in a chemical fume hood.
4. To each dish containing cell culture media and formaldehyde, add Glycine Solution (10X) to a final concentration of 1X. Mix well by gently swirling the dish. Incubate at room temperature for 5 minutes in the chemical fume hood.
5. Aspirate formaldehyde/glycine-containing media in the fume hood. Properly dispose the formaldehyde-containing waste.
6. Wash the cells twice with one media volume of ice-cold PBS, removing each wash by aspiration.
7. Add 10 μ L of the Halt Cocktail to 1mL of ice-cold PBS. Add the solution to the cells, and detach cells by scraping. Transfer the cell suspension to a 1.5mL microcentrifuge tube using a pipette
8. Centrifuge tubes at 3000 \times g for 5 minutes.
9. Remove the PBS. Store the cell pellet(s) at -80°C, or proceed directly to Section D: Lysis and MNase Digestion.

D. Lysis and MNase Digestion

Note: For best results, empirically determine the optimal crosslinking time and Micrococcal Nuclease digestion conditions for each cell type (see Appendix A).

1. Use the crosslinked cells prepared above. If frozen, thaw cells on ice.
2. Add 100 μ L Lysis Buffer 1 containing protease inhibitors to the cell pellet and pipette up and down to break up the pellet. Vortex the tube for 15 seconds and incubate on ice for 10 minutes.
3. Centrifuge at 9000 \times g for 3 minutes and remove the supernatant.
4. Resuspend nuclei in 100 μ L of MNase Digestion Buffer Working Solution.
5. Add 0.25 μ L of Micrococcal Nuclease (CHIP Grade) (10 U/ μ L), vortex the tube and incubate in a 37°C water bath for 15 minutes, mixing by inversion every 5 minutes.
6. Add 10 μ L of MNase Stop Solution to stop the reaction, vortex briefly and incubate on ice for 5 minutes.
7. Centrifuge at 9000 \times g for 5 minutes to recover the nuclei. Remove the supernatant.
8. Resuspend nuclei in 50 μ L of Lysis Buffer 2 containing protease/phosphatase inhibitors and incubate on ice for 15 minutes, vortexing for 15 seconds every 5 minutes.
9. Centrifuge at 9000 \times g for 5 minutes and transfer the supernatant, containing the digested chromatin, to a new 1.5mL tube. Proceed to the immunoprecipitation or store samples at -80°C.

E. Immunoprecipitation

1. Transfer 5 μ L of the supernatant containing the digested chromatin to a 1.5mL tube and store at -20°C. This is the 10% total input sample from one CHIP.

Note: If you are preparing chromatin in bulk, this sample can be used as a common input control for all IPs from the bulk preparation.

2. Transfer the remaining 45 μ L of supernatant to 450 μ L of 1X IP Dilution Buffer.

Note: If you are preparing chromatin in bulk, unused supernatant may be stored at -80°C for later use.

-
- For each IP, add 500 μ L diluted lysate to a plugged spin column and add primary antibody. The following amounts of antibody are recommended when using 2×10^6 HeLa cells/IP.
 - Positive control IP (included): add 10 μ L Anti-RNA Polymerase II Antibody
 - Negative control IP (included): add 1 to 2 μ L of Normal Rabbit IgG
 - Target-specific IP(s): The typical concentration is 1-10 μ g antibody for each IP; however, titration of antibody concentration to obtain the best signal-to-noise ratio is required.
 - Incubate IP reactions for 2 hours to overnight at 4 $^{\circ}$ C on a rocking platform.

Note: For low-abundant proteins, an overnight incubation greatly increases signal.
 - Cap and vortex the tube of ChIP Grade Protein A/G Plus Agarose to obtain a uniform suspension. Using a wide-bore or cut pipette tip, add 20 μ L of the agarose resin to each IP and incubate for 1 hour at 4 $^{\circ}$ C on a rocking platform.
 - After resin incubation, remove and set aside the column plug. Place the column into a 2mL collection tube.
 - Centrifuge tube at $3000 \times g$ for 30 seconds. Remove column from the collection tube, discard the column flow-through and set aside the collection tube.
 - Re-insert the same column plug, apply 0.5mL of IP Wash Buffer 1, cap the column and incubate at 4 $^{\circ}$ C for 5 minutes on a rocking platform.
 - Remove the column plug and set aside. Place the column into the same collection tube.
 - Centrifuge at $3000 \times g$ for 30 seconds. Remove column from the collection tube, discard the column flow-through and set aside the collection tube.
 - Repeat steps 8-10 twice with IP Wash Buffer 2.
 - Repeat steps 8-10 once with IP Wash Buffer 3.
 - To remove residual Wash Buffer, insert column into the collection tube and centrifuge at $3000 \times g$ for 1 minute.

F. IP Elution

- Re-fit column plug, add 150 μ L 1X IP Elution Buffer to the washed resin, cap the column, and incubate at 65 $^{\circ}$ C for 30 minutes with shaking. If a thermomixer is unavailable, incubate the column in a heat block set a 65 $^{\circ}$ C for 40 minutes. Re-suspend the beads by flicking the tube every 10 minutes.
- During the elution step, prepare a 1.5mL microcentrifuge tube for each IP containing 6 μ L of 5M NaCl and 2 μ L of 20mg/mL Proteinase K.
- Thaw 10% total input sample(s) and add 150 μ L of IP Elution buffer, 6 μ L of 5M NaCl and 2 μ L of 20mg/mL Proteinase K. Place tubes at room temperature until the next step.
- Following the 65 $^{\circ}$ C incubation, remove the column from the heat block, open the cap, remove the plug, and place the column in the previously prepared 1.5mL collection tube containing the NaCl and Proteinase K, close the column cap and centrifuge at $6000 \times g$ for 1 minute.

Note: After incubation, it is essential to open the column cap before plug removal to equalize the pressure within the column. Failure to open the cap before plug removal will result in sample loss.
- Discard the columns. Cap the 1.5mL centrifuge tubes, vortex all IP and total-input samples and place in a 65 $^{\circ}$ C heat block for 1.5 hours.

G. DNA Recovery

- To each eluted IP and total input sample, add 750 μ L of DNA Binding Buffer.
- Pipette 500 μ L of each sample into a DNA Clean-Up Column inserted into a 2mL collection tube. Centrifuge the columns at $10,000 \times g$ for 1 minute and discard the flow-through.
- Pipette the remaining sample into the same DNA Clean-Up Column. Centrifuge the columns at $10,000 \times g$ for 1 minute and discard the flow-through.
- Place the column back into the collection tube and add 750 μ L of DNA Column Wash Buffer. Centrifuge the columns at $10,000 \times g$ for 1 minute and discard the flow-through.

5. Place the column back in the empty collection tube and centrifuge the columns at $10,000 \times g$ for 2 minutes.
6. Place the column in a new 1.5mL centrifuge tube and pipette 50 μ L of DNA Column Elution Solution directly into the center of each column.
7. Centrifuge the column at $10,000 \times g$ for 1 minute and discard the column. The resulting solution is the purified DNA. Proceed to PCR or QPCR detection (see Appendix B).

Troubleshooting

Problem	Possible Cause	Solution
Chromatin fragments are too large (> 1000 base pairs)	Crosslinking time was too long	Shorten crosslinking time
	Cell to Micrococcal Nuclease (MNase) ratio was too high	Increase amount of MNase or decrease cell number (refer to the MNase digestion optimization protocol in Appendix A)
Chromatin fragments are too small (< 200 base pairs)	Cell to MNase ratio was too low	Decrease amount of MNase or increase cell number (refer to the MNase digestion optimization protocol in Appendix A)
Chromatin not recovered from nucleus	Cell type requires more stringent handling to better lyse and release chromatin from nucleus	Increase incubation with nuclei lysis buffer to 30 minutes and vortex for 30 seconds every 5 minutes
		Following the nuclear lysis step, sonicate the sample for 60 seconds using a 1/8" probe. Perform on wet ice in 3 pulses of 20 seconds with 30 second pauses between
		Following the nuclear lysis step, dounce the sample 20 times in a glass dounce homogenizer
No or low PCR signal in the total input control samples	PCR amplification conditions were not fully optimized	Optimize PCR conditions using samples known to contain the target amplicon
		Check primer design
	Insufficient amount of sample DNA added to the PCR reaction	Increase the amount of sample DNA added to the PCR reaction
No or low PCR signal in the positive control IP samples	Insufficient chromatin amount in the IP reaction	Use at least 25 μ g of chromatin for each IP
	Insufficient antibody incubation time	Incubate antibody overnight
	Incomplete elution from the Protein A/G agarose resin	Perform elution at 65°C and increase frequency of mixing
PCR signal of the positive and negative control IP samples are equivalent	Insufficient washing of the IP complex	Include an additional wash with Buffers 2 and 3
	Excess chromatin or antibody added to the IP	Add less chromatin or antibody
	PCR amplification was measured outside the linear range of amplification	Decrease the number of amplification cycles used in the PCR reaction

Continued on next page

Continued from previous page

No or low PCR signal in the experimental IP samples	Insufficient amount of sample DNA added to the PCR reaction	Add more sample DNA to the PCR reaction
	Insufficient amount of antibody added to the IP	Add more antibody to the IP
	Antibody does not function in an IP	Verify that the antibody is qualified for ChIP or IP applications

Note: Increasing the stringency of the nuclear lysis is not recommended unless there is no signal or low signal-to-noise in the IP. Excess sample can result in high background, decreasing the signal of the specific signal.

Appendix A: Optimization of micrococcal nuclease digestion

1. Prepare crosslinked cell pellet from 10^7 cells as described in Section C.
2. Follow the lysis procedure in Section D, Steps 1-4; however, use 0.5mL of the Lysis Buffer 1, Membrane Extraction Buffer and MNase Digestion Buffer Working Solution.
3. Transfer 100 μ L of the resuspended nuclei into five tubes (2×10^6 nuclei/tube), labeling the tubes 0, 1, 2, 4 and 6.
4. Add 2 μ L of the Micrococcal Nuclease (ChIP Grade) to 18 μ L of MNase Digestion Buffer Working Solution (1:10 dilution).
5. Add 0, 1, 2, 4 and 6 μ L of the diluted Micrococcal Nuclease to the corresponding tubes of nuclei and incubate for 15 minutes in a 37°C water bath mixing by inversion every 5 minutes.
6. Add 20 μ L MNase Stop Solution to each tube and incubate on ice for 5 minutes.
7. Centrifuge at $9000 \times g$ for 5 minutes to recover the nuclei and remove the supernatant.
8. Resuspend nuclei in 50 μ L of Nuclear Extraction Buffer and incubate on ice for 15 minutes, vortexing for 15 seconds every 5 minutes.
9. Centrifuge at $9000 \times g$ for 5 minutes.
10. Prepare five 1.5mL microcentrifuge tubes labeled 0, 2, 4 and 6 each containing 6.6 μ L of Nuclease Free Water, 2.4 μ L of 5M NaCl, and 1 μ L of Proteinase K (20mg/mL).
11. Transfer 50 μ L of each nuclear supernatant, containing digested chromatin, to the corresponding tube.
12. Vortex each tube for 10 seconds and incubate the tubes at 65°C in a heat block for 1.5 hours.
13. Analyze 15-25 μ L of each sample and DNA size marker by agarose gel electrophoresis. Optimal digestion should yield fragments from 200 to 1000 base pairs with a more intense ladder of bands occurring at approximately 160, 320, and 480 base pairs, which corresponds to the 1, 2 and 3 nucleosome units.

Appendix B: Real-time PCR analysis using positive control primers

Recommendations:

- Use filter tips, gloves, and nuclease-free reagents to prevent contamination.
- Using hot-start Taq polymerase and validated PCR master mix greatly reduces PCR artifacts and increases reaction efficiency.
- The Positive Control Primers included are designed to amplify a region of the human GAPDH promoter close to the transcription start site. The primers are validated for detecting RNA polymerase II binding to the GAPDH promoter in human-derived cells and tissue. Detection of other Protein:DNA interactions using these primers is not validated.
- When designing primers, using reputable primer design software greatly increases successful PCR detection.
- Follow the manufacturer's recommendations when programming the thermocycler and collecting real-time data.
- Use a standard curve of serially diluted genomic DNA to evaluate PCR efficiency and linear amplification.
- Amplify each sample in triplicate to control for pipetting error.

Recommended reaction conditions for quantitative real-time PCR (qPCR) with the GAPDH control:

Reagent	Volume/reaction (μL)
Thermo Scientific™ 2X Absolute QPCR SYBR™ Green Fluorescein Mix	12.5
ChIP Positive Control Primers (GAPDH Promoter)	1.0
Template DNA	5.0
Nuclease-free Water	6.5

Thermocycler Amplification Settings:

Step 1: 95°C for 15 minutes.

Step 2: 95°C for 15 seconds.

Step 3: 62°C for 1 minute (collect real-time data).

Step 4: Repeat Steps 2 to 3 for 40 cycles.

Related Thermo Scientific Products

See the website for a listing of ChIP-validated antibodies.

28908	16% Formaldehyde (w/v), Methanol-free, 10 × 1mL
20688	Dimethylsulfoxide (DMSO), Sequanal grade, 950mL
26158	Pierce Chromatin Prep Module
26159	ChIP Grade Protein A/G Plus, 0.65mL
26160	Proteinase K Solution (20mg/mL), 0.25mL
28372	BupH™ Phosphate Buffered Saline Pack, 40 packs
78425	Halt Protease Inhibitor Cocktail, EDTA-Free (100X), 24 vials
78443	Halt Protease and Phosphatase Inhibitor Cocktail, EDTA-Free (100X), 24 vials
78833	NE-PER Nuclear and Cytoplasmic Extraction Kit
20148	LightShift™ Chemiluminescent EMSA Kit
23225	Pierce BCA Protein Assay Kit
22662	Pierce 660nm Protein Assay Kit
AB1219	Absolute™ QPCR SYBR™ Green Fluorescein Mix
AB-0900/w	Thermo-Fast™ 96 Semi-Skirted PCR Plate, opaque white
AB-1170	ABsolute QPCR Seal

Cited References

- Lee, D.Y., *et al.* (2009). Role of protein methylation in regulation of transcription. *Endocrine Rev* **26(2)**:147-70.
- Zeng, P.Y., *et al.* (2006). *In vivo* dual cross-linking for identification of indirect DNA-associated proteins by chromatin immunoprecipitation. *Biotechniques* **41**:694-8.
- Aparicio, O., *et al.* (2004). Chromatin immunoprecipitation for determining the association of proteins with specific genomic sequences *in vivo*. *Curr Protoc Cell Biol* Chapter 17:Unit **17.7**.

General References

- Das, P.M., *et al.* (2004). Chromatin immunoprecipitation assay. *Biotechniques* **6**:961-9.
- Wells, J. and Farnham, P.J. (2002). Characterizing transcription factor binding sites using formaldehyde crosslinking and immunoprecipitation. *Methods* **1**:48-56.
- Yuichi Y. and Hughes D.E., (2006). C/EBPα and HNF6 protein complex formation stimulates HNF6-dependent transcription by CBP coactivator recruitment in HepG2 cells. *Hepatology* **43(2)**:276-86.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts (“Documentation”). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample.

NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON INFRINGEMENT. BUYER’S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER’S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS.

Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals.

Current product instructions are available at thermofisher.com. For a faxed copy, call 800-874-3723 or contact your local distributor.

© 2016 Thermo Fisher Scientific Inc. All rights reserved. Unless otherwise indicated, all trademarks are property of Thermo Fisher Scientific Inc. and its subsidiaries. Printed in the USA.