

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 4519-110

**Product Description:** Insertion Probe Fitting, Vertical Sidearm Flask, pH Probe

**Component Materials:**

Cap - Polypropylene (PP)  
Nut - Polyphenylene Oxide (PPO) / Polystyrene (PS)  
Insert - Polyethylene terephthalate (PET)  
O-ring - Synthetic Rubber

**Attributes:**

Product Dimensions - 54 mm OD x 52 mm High

**Autoclaving:**

This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

**Heating / Cooling:**

This product is primarily designed for a specific use that does not involve direct heating. In normal use, temperatures will be from -50° to +125°C. This is below the temperature at which thermal failure could be a problem.

**Vacuum / Pressure:**

This product is not designed for and is not rated for vacuum or pressure applications.

**Sterilization:**

Shipped Not Sterile

**Lot Number Designation:**

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Revision Date:**

5 AUG 2013

Rev No: 1