Corning Incorporated Life Sciences Registered ISO 9001:2008		Product	Product Description	
Catalog Number:	3736			
Product Description:	Costar ® 6-v	well non-treated cell cultu	re multiple well plate,	flat bottom, with lid
Component Materials: Plate/Lid -	Virgin Polystyrene, meets USP, Class VI requirements for plastics containers and closures.			
<b>Product Dimensions:</b> Length of Plate Width of Plate	- 5.030 in. - 3.365 in.		er of Well @ top - er of Well @ -	1.395 in. 1.370 in.
Depth of Well Tolerances of Dimensions	685 in. - +/050 ir	Height v	vith Lid - 9 Volume per well -	.891 in. 5 mL

# Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10<sup>-3</sup>

### Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing. Results:  $\leq 0.1 \text{ EU/mL}$  ( $\leq 4 \text{EU/device}$ )

# Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured with animal free materials.

# Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. No Surface Treatment: Wettability test using water to insure the presence of a hydrophobic

Visual Attributes:surface.Packaging:Visual and microscopic examination of the product.Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

# Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5