

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3383

Product Description: Corning® HTS Transwell, Feeder Reservoir, with lid, with Liquid Media Stabilizer

Component Materials:

- Reservoir & Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
Liquid Media Stabilizer - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|--------------------------|---|--------------|--------------------|---|----------|
| Length of reservoir | - | 5.030 in. | Depth of Reservoir | - | .442 in. |
| Width of Reservoir | - | 3.365 in. | Height with Lid | - | .626 in. |
| Tolerances of Dimensions | - | +/- .010 in. | | | |

Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4