## Corning Incorporated Life Sciences

Registered ISO 9001

# **Product Description**

Catalog Number: 8388

Product Description: Costar ® 125 mL Storage Bottle, plug seal cap, 45 mm neck

**Component Materials:** 

Bottle - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Cap - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic

containers and closures. Heavy metal free (meets CONEG reg.) color concentrate.

**Product Dimensions:** 

Height without cap - 4.02 in. O.D. (neck) incl. Threads - 1.74 in. Height of bottle w/cap - 4.07 in. I.D. (neck) - 1.52 in.

Tolerances - +/- 0.05

#### Sterilization:

The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10<sup>-3</sup>

#### Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing. Results: less than 0.1EU/mL.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct

product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Run number for that year.

### **Revision Date:**

10-17-18

Rev No: 9