

**Catalog Number:** 3358

**Product Description:** Costar® 96-well cell culture cluster, round bottom, with lid

**Component Materials:**

Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

Length of plate	-	5.030 in.	Diameter of well @ top	-	0.270 in.
Width of Plate	-	3.365 in.	Diameter of well @ bottom	-	0.250 in.
Depth to the bottom of well	-	0.445 in.	Height with lid	-	0.650 in.
Tolerances of dimensions	-	+/- 0.010 in.	Working volume per well	-	0.36mL

**Sterility-** Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of  $10^{-3}$

**Non-Pyrogenic-** Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is  $\leq 0.10$  EU/ml or  $\leq 4$  EU/device.

**Animal Content** - Product does not contain materials of animal origin.

**Surface Characterization** – Surface is characterized to be hydrophilic and negatively charged, composed of 9-17% oxygen atoms. This surface composition has been optimized for cell attachment and growth.

**Tissue Culture** - Tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line. A minimum of 95% confluency is required for acceptance.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass

Packaging Inspection – Pass

Cell Attachment & Growth treatment verification – Pass

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture, Last 3 digits – Batch identification.