## Corning Incorporated Life Sciences

Registered ISO 9001:2008 **Product Description** 

Catalog Number:		3382			
Product Description:		Corning ® 96-well HTS Receiver Plate, with low evaporation lid			
<b>Component Materials:</b> Plate/Lid	-	Virgin Polystyrene meets UPS, Class VI requirements for plastic containers and closures.			
<b>Product Dimensions:</b> Length of Plate Width of Plate Tolerances of Dimensions	- -	5.030 in. 3.365 in. +/010 in.	Depth of the bottom of well Height with Lid Working Volume per well	- -	.446 in. .630 in. 300 μL

Animal Content: - Product does not contain materials of animal origin.

**Sterility:** – Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137. "Sterilization of health care products – Radiation". Products meet a minimum of Sterility Assurance Level (SAL) of 10<sup>-3</sup>.

**Tissue Culture**: - Tested for the attribute of cell attachment inhibition utilizing an attachment- dependent mammalian cell line. A minimum of 95% cell attachment inhibition is required for acceptance.

**Optical Characterization:** – The product is made of clear polystyrene to allow transparency.

**Quality Control Testing:** – Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below: Visual Inspection - Pass Packaging Inspection – Pass

## Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.

Rev 5