

Genexus™ GX5™ Starter Pack-AS

Catalog Number A40279

Pub. No. MAN0017913 Rev. B.0

Note: For safety and biohazard guidelines, see the “Safety” appendix in the *Genexus™ Integrated Sequencer User Guide* (Pub. No. MAN0017910). Read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

Product description

The Ion Torrent™ Genexus™ GX5™ Starter Pack-AS for Ion AmpliSeq™ panels provides one set of all core consumables required to run the Genexus™ Integrated Sequencer. This starter pack supports flexible, fully automated preparation and normalization for up to 32 Ion AmpliSeq™ libraries, as well as sequencing for up to eight chip lanes (two full chips). The GX5™ Chip and Genexus™ Coupler (Cat. No. A40269) are used in combination with this starter pack, and must be ordered separately.

Note: Genexus™ Strip 1, Genexus™ Strip 2-AS, Genexus™ Strip 3-GX5™, and Genexus™ Strip 4 with part numbers listed in the following table are compatible only with Genexus™ Software 6.2.0 and later.

Contents and storage

Component	Part or Cat. No.	Quantity	Storage
Genexus™ Strip 1	A46812	8 strips	2°C to 8°C
Genexus™ Strip 2-AS	A46813	8 strips	-30°C to -10°C
Genexus™ Strip 3-GX5™	A46815	8 strips	2°C to 8°C
Genexus™ Strip 4	A46816	8 strips	-30°C to -10°C
Genexus™ Barcodes 1–32 AS	A40258	1 plate	15°C to 30°C
Genexus™ Pipette Tips	A40266	12 racks	
Genexus™ Cartridge	A40272	2 cartridges	-30°C to -10°C
Genexus™ Bottle 2	A40273	4 bottles	15°C to 30°C
Genexus™ Bottles 1 and 3	A40274	2 bottles each	

Note:

- Genexus™ Barcodes 1–32 AS and Genexus™ Pipette Tips are also orderable individually using the catalog numbers listed. Genexus™ Barcodes 33–64 AS (Cat. No. A40259), Genexus™ Barcodes 65–96 AS (Cat. No. A40260), and Genexus™ Barcodes 1–96 AS (Cat. No. A40257) are also orderable individually.
- Library and template strips are orderable separately in paired sets as Genexus™ Library Strips 1 and 2-AS (Cat. No. A40252) and Genexus™ Templating Strips 3 and 4-GX5™ (Cat. No. A40263).
- The Genexus™ Cartridge, Genexus™ Bottle 2, and Genexus™ Bottles 1 and 3 are orderable as a set in the Genexus™ Sequencing Kit (Cat. No. A40271).
- For custom assays, Genexus™ Primer Pool Tubes (Cat. No. A40262) must be ordered separately.

Precautions

Handle biological samples and all other materials according to local laboratory guidelines and regulations. Avoid cross-contamination of samples and reagents to ensure quality results.

If equipment or reagents are used in a manner not specified by the manufacturer, the protection provided by the equipment can be impaired.

IMPORTANT! Do not substitute components in this kit with components from other Ion Torrent™ templating or sequencing kits. This protocol has been verified using these specific materials. Substitution can adversely affect system performance.

Note: Refer to the expiration date on the package label. Do not use products after the expiration date.

Customer and technical support

Visit thermofisher.com/support for the latest service and support information.

- Worldwide contact telephone numbers
- Product support information
 - Product FAQs
 - Software, patches, and updates
 - Training for many applications and instruments
- Order and web support
- Product documentation
 - User guides, manuals, and protocols
 - Certificates of Analysis
 - Safety Data Sheets (SDSs; also known as MSDSs)

Note: For SDSs for reagents and chemicals from other manufacturers, contact the manufacturer.

Limited product warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale at www.thermofisher.com/us/en/home/global/terms-and-conditions.html. If you have any questions, please contact Life Technologies at www.thermofisher.com/support.



Life Technologies Corporation | 7335 Executive Way | Frederick, MD 21704 | USA

For descriptions of symbols on product labels or product documents, go to thermofisher.com/symbols-definition.

The information in this guide is subject to change without notice.

DISCLAIMER: TO THE EXTENT ALLOWED BY LAW, THERMO FISHER SCIENTIFIC INC. AND/OR ITS AFFILIATE(S) WILL NOT BE LIABLE FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE, OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING YOUR USE OF IT.

Revision history: Pub. No. MAN0017913

Revision	Date	Description
B.0	7 July 2020	Updated Genexus™ reagent strip part numbers
A.0	11 November 2019	New product information sheet for the Genexus™ GX5™ Starter Pack-AS

Important Licensing Information: This product may be covered by one or more Limited Use Label Licenses. By use of this product, you accept the terms and conditions of all applicable Limited Use Label Licenses.

©2020 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.